

563 Shoreview Park Road St. Paul, MN 55126 Phone 651-484-6544 Fax 651-484-6824

# **EC Declaration of Conformity**

MANUFACTURER: Lion Precision

563 Shoreview Park Road St. Paul, MN 55126-7014

APPARATUS: Eddy Current Sensor

**TYPES**: *ECL100 & ECL101* 

This Apparatus conforms to the following European Directives:

Council Directive 2004/108/EC Electromagnetic Compatibility Directive.

The following standards were used to assess the products:

**Standard:** IEC 61326-1:2005-12

Electrical equipment for measurement, control and laboratory use EMC

requirements.

**Standard:** IEC 61326-2-3:2006-07

Particular Requirements – Test Configuration, operational conditions and

performance criteria for transducers with integrated or remote signal

conditioning..

Standard: EN 61010-1:2001 IEC 61010-1:2001(+C1 & C2)

Safety requirements for electrical equipment for measurement, control, and

laboratory use.

Date of issue: June 26, 2009

Don Martin, President

## **IMMUNITY: IEC 61326-2-3**

## **ENCLOSURE:**

TEST	STANDARD	TEST VALUE	COMPLIANCE
ELECTROSTATIC DISCHARGE	IEC 61000-4-2	4 kV/8 kV contact/air	Criterion A
EM FIELD	IEC 61000-4-3	10 V/m	Criterion A
RADIATED POWER FREQ MAGNETIC FIELD	IEC 61000-4-8	30 A/m	Criterion A

#### **AC POWER:**

TEST	STANDARD	TEST VALUE	COMPLIANCE
VOLTAGE DIP/SHORT INTERRUPTIONS	IEC 61000-4-11	0,5 cycle, each polarity/100%	Criterion A
BURST	IEC 61000-4-4	2 kV	Criterion A
SURGE	IEC 61000-4-5	1 kV/2 kV	Criterion A
CONDUCTED RF	IEC 61000-4-6	3 V	Criterion A

## I/O SIGNAL/CONTROL:

TEST	STANDARD	TEST VALUE	COMPLIANCE
BURST	IEC 61000-4-4	1 kV	Criterion A
SURGE	IEC 61000-4-5	1 kV	N/A (<30m)
CONDUCTED RF	IEC 61000-4-6	3 V	Criterion B

## EMISSIONS: IEC 61326-2-3 CLASS A:

TEST	RANGE	LIMITS
ENCLOSURE	30 to 230MHz	40db (uV/M)quasi peak, measured at 10 m
	230 to 1000MHz	47db (uV/M)quasi peak, measured at 10 m
AC MAINS	0.15 to 0.5 MHz	79db (uV)quasi peak
		66db average
	0.5 to 5 MHz	73db (uV)quasi peak
		60db average
	5 to 30 MHz	73db (uV)quasi peak
		60db average

## ELECTRICAL SAFETY: EN 61010-1:2001

A representative unit has been inspected and tested against the relevant requirements of BS EN 61010-1.